



Mascarilla quirúrgica 3 capas 3 pliegues. Protección BFE-VFE-PFE mayor/igual a 95%. Cada mascarilla en bolsa individual. Capa exterior color celeste.



# **DESCRIPCIÓN**

Mascarillas desechable plana con tiras de sujeción, posee una eficiencia de filtración bacteriana de 99%. Protege la nariz y la boca para minimizar los contactos infecciosos en los medios quirúrgicos, dental, médico y en todos los demás ambientes estériles o sensibles. Esta fabricada con 3 capas de tela no tejida (capa externa 25 g / fabric tela no tejida doble S, capa interna 25 g / m? tela no tejida respetuosa con la piel), una capa de tela soplada por fusión (Eficiencia de filtración> 95%, 25 g / m?), la capacidad de filtración general supera el 95%.

## Características

MATERIAL: Tela no tejida de plipropileno

Número de capas: 3

Capa interna: Tela no tejida de polipropileno, tela suave.

Capa media (filtro)

Capa externa: Tela no tejida de polipropileno, a prueba de agua.

#### Diseño:

Mascarilla descartable de 3 pliegues con tiras de sujeción.

Debe estar dotada de un medio por el cual se puede ceñir estrechamente sobre la nariz, boca y barbilla de quien la lleva puesta y que garantice que la mascarilla se ciñe estrechamente en los laterales.

Puente nasal: Barra moldeable de adaptación anatómica a la nariz en el borde superior de la mascarilla, la misma que debe estar protegida para no lesionar la pielm ni producir incomodidad al usuario.

Requisitos de desempeño para las mascarillas quirúrgicas Tipo I

Eficacia de filtración bacteriana (BFE) (%): >=95

Presión Diferencial (Pa/cm2): <40

Carga Biológica (limpieza microbiana) (ufc/g): <=30

### Marcado, etiquetado y envasado

Tiene la siguiente información indicada en el envase en que se suministra la mascarilla:

- a) Número de la norma técnica peruana (NTP 329.200), mediante sello o sticker.
- b) Presentación en cajas de 50 unidades.







## **Declaration of Conformity**

Manufacturer: Anqing Mayfield Medical Limited

Jingwu Road, Tongcheng Economic and Technological Development

Zone, Anqing City, Anhui Province, CHINA

whose single Luxus Lebenswelt GmbH

Authorized EU- Kochstr.1, 47877, Willich, Germany

Representative: DIMID: DE/0000047791

Lin Sun

Tel: 0049- 1715605732 E-mail: info.m@luxuslw.de

Product Name: Disposable Face Masks

Classification: I of ANNEX IX of Directive 93/42/EEC

Conformity Assessment Route: Annex III

We here declared that above mentioned product met the transposition into national law, the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.

General Applicable Directives: Medical Devices DIRECTIVE

Harmonized Standards:

EN ISO 13485:2016

EN ISO 10993

EN 980-2003

EN 14683

EN ISO 14971-2007

Signature:

Name: 周苹

Title: +86-0556-6181088

Position: General manager

Sea 出 2020-07 7 3-08 8 10 1 5 8 8 2 1





### Fiscal Year 2020

### **VERIFICATION OF FDA REGISTRATION**

This certifies that:

### Anqing Mayfield Medical Limited

Jingwu Road, Tongcheng Economic and Technological Development Zone, Anqing City, Anhui Province, 231400, CHINA

has completed the FDA Establishment Registration and Device Listing with the US Food & Drug Administration, through UCL-REG SERVICE INC.

Owner/Operator Number: 10076152

Listing No. Product Code:

D411432 QKR Face mask (except N95 respirator) for general public/healthcare personnel per IIE guidance-

Disposable Face Mask
Respirator, surgicalFolding Mask

**Device Name:** 

D411433 MSH

UCL-REG SERVICE INC. will confirm that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. UCL-REG SERVICE INC. makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. UCL-REG SERVICE INC. assumes no liability to any person or entity in connection with the foregoing.

Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding."

The U.S.Food and Drug Administration does not issue a certificate of registration, nor does the U.S Food and Drug Administration recognize a certificate of registration, UCL-REG SERVICE INC. is not affiliated with the U.S. Food and Drug Administration.



UCL-REGSERVICE INC.
602 ROCKWOOD ROAD, WILMINGTON,
NEW CASTLE DE 19802 USA

UCL-REG SERVICE INC

Cert. No.: M201061 Issued Date: July 07 2020 Expiration Date: 31 December 2021